

Cone Beam CT dosimetrie

Technique	Effective Dose (μSv)	Compared to the dose from panoramic radiograph	Percentage of the annual background dose
Intraoraal	5	0.2-0.25	0.1%
OPT	22	1	0.6%
Cone Beam CT	50-600	3-30	5%
Medical CT	400-1200	20-60	40%

Bron: SedentexCT, provisional guidelines, 2009

